

ASSAM CANCER CARE FOUNDATION (ACCF)

3rd floor, V.K. Trade Centre, G.S. Road, Opp. Down-town Hospital, Guwahati – 781022, Assam Ph: +91-90852 02020

E: procurement@accf.in | W: www.assamcancercarefoundation.org

For supply, installation & commissioning of medical equipment, surgical instrument and hospital furniture

Ref: ACCF/Medical Equipment/2021-22/30

Corrigendum-1(PRE-BID RESOLUTION)

Date: 09.12.2021

Sr.No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution/Amendment(s)
1	7.2 Quality Standard	7.2.3 Sample can be asked from the bidder and approved sample will be retained by TIA to cross check the quality at the time of delivery.	Clarification	Samples of Surgical Instruments (small, medium and large) are required to be submitted by bidders at the time of bid submission or within 5 days from the date of bid submission date. Regarding Furniture and medical equipment, date of demonstration shall be intimated by ACCF separately.
2	Group E Sr. No. 76	Nasopharyngoscope		Item Deleted
3	4.2.8	4.2.8 Payment Timeline 70% Payment shall be released against successful Delivery. 20% shall be released on successful installation, subject to verified installation report and rest 10 % after final acceptance certificate of completion from the user/Site Engineer/BME		Payment terms: 1) For items which do not require installation payment terms shall be: 90% Payment shall be released against successful acceptance of goods. 10% shall be released after 30 days of acceptance by the user/Site Engineer/BME. 2) For items which require installation payment terms shall be: 70% Payment shall be released against successful delivery. 20% shall be released on successful installation, subject to verified installation report and rest 10 % after final acceptance certificate of completion from the user/Site Engineer/BME
4	GROUP E SR No.73	Compatible 300-Watt Xenon light source with coloured temperature around 6000 kelvin and Led lamp as auxiliary / back up. Automatic light adjustment to maintain optimum brightness. It should have a coloured system CCD 2 spare bulbs (same quality) It should be compatible to all scopes and ULTRASOUND endoscope and transmit image digitally It should have automatic as well as manual brightness control mode. It should have facility of extra illumination for more light apart from brightness control. Processor should be able to give images of surface analysis and vessel analysis for identifying lesions and perform improve pit pattern classification.	(ii) HD Video Processor and Cold Light Source (Both from Original Manufacturer} Page No. 63 it is mentioned Compatible 300 watt Xenon Light Source and LED Lamp whereas in Page 64 under Specification Required Point No. 7, it is mentioned Light Source Xenon or LED Light at least 300 W.** Kindly amend it to 300 Watt Xenon/ LED	Compatible 300-Watt Xenon/LED light source with coloured temperature around 6000 kelvin and Led lamp as auxiliary / back up. Automatic light adjustment to maintain optimum brightness. It should have a coloured system CCD 2 spare bulbs (same quality) It should be compatible to all scopes and ULTRASOUND endoscope and transmit image digitally It should have automatic as well as manual brightness control mode. It should have facility of extra illumination for more light apart from brightness control. Processor should be able to give images of surface analysis and vessel analysis for identifying lesions and perform improve pit pattern classification.
5	GROUP E SR No.73 Pt.02	It should be compatible to all scopes and	It is mentioned that "It should be compatible with Ultrasound	No Change

Sr.No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution/Amendment(s)
		ULTRASOUND endoscope and transmit image digitally	Endoscope *** Future compatibility to EUS System results in high end and expensive system to be quoted, therefore kindly make is desirable.	
6	SECTION-III EMD	GROUP E Sr. No.73: 100,000.00 Sr. No.74: 100,000.00 Sr. No.75: 100,000.00 Sr. No.76: 100,000.00	Earnest Money Deposit In Page No. 4-The EMO Amount is Rs. 50,000/- for Sr.No. 74 In Page No. 23 - The EMO Amount is Rs. 1,00,000/- for Sr. No. 74** Kindly confirm, which is correct.	GROUP E Sr. No.73: 50,000.00 Sr. No.74: 50,000.00 Sr. No.75: 50,000.00 Sr. No.76: 50,000.00
7	CAMC	Page 24:(warranty 2 years from the date of Commissioning, 5 years CMC after warranty) Page 63(Bronchoscope, Adult: -Warranty: 5 years after complete installation, CAMC for 5 years after expiry of warranty period. CAMC should also include.) Page 64(Two years warranty with five years AMC/CMC. Approval of the equipment would be subject to the working demonstration as per ACCF descretion)	Comprehensive Warranty/ CAMC In Page No. 24- It is mentioned 2 Years Warranty & 5 Years CMC/AMC In Page No. 63 - It is mentioned 5 Years Warranty & 5 Years CMC/AMC In Page No. 64 -- It is mentioned 2 Years Warranty & 5 Years CMC/AMC ** Kindly confirm which is correct.	2 years Warranty and post warranty 5 year CMC
8	Page 63. Scope Set - Bronchoscope, Adult:	Pg. 63 Scope Set -Bronchoscope, Adult:Field view:1100 or More,Bending Angulation range:Up-1800 -Down-1200 or more	Specification required under Adult Bronchoscope Field of View - 1100 or more - It should be 110 Degree or more Bending Angulation Range - Up 1800 or Down 1200 or more - It should be UP 180 Deg and Down 120 Deg or rmore	Pg. 63 Scope Set -Bronchoscope, Adult:Field view:110 degree or More,Bending Angulation range:Up-180 degree -Down-120 degree or more
9	Page 64.Scope Set - Bronchoscope, Pediatric:	Pg. 64 Scope Set -Bronchoscope, Pediatric: Field of view: at least 1000 Bending Section Angulations:range UP 1800, DOWN 1300	Same mistake found in Specification required under Paediatric Bronchoscope	Pg. 64 Scope Set -Bronchoscope, Pediatric: Field of view: at least 100 degree Bending Section Angulations:range UP 180 degree, DOWN 130 degree
10	Page 63. Scope Set - Bronchoscope, Adult: Certifications	All the material /equipment should be European CE/ US FDA Certified.	Certification Required - It should have European CE/ US FDA We requested you to include ICMED the Indian Certification for Medical Device. ICMED is India's first indigenously developed International Class assurance system for Indian manufacturer medical device,We wish to inform you that USFDA is not mandatory for India and is against the spirit of "Atmanirbhar Bharat". Please find attached the directive issued by Government of India in this regards.We now request you to kindly amend the Tender Clause	No change
11	BP Apparatus - Dial	BP Apparatus - Dial Cuff - Velcro cuffs or equivalent fastening design. Adult size 54,5 x 14,5 cm Small adult size 42 x 13 cm, 1 tube, 2 tubes	Velro cuffs For better participation	No change

Sr.No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution/Amendment(s)
		Obese size 70 x 15 cm, 1 tube, 2 tubes Thigh size 70 x 22 cm, 1 tube, 2 tubes		
12	BP Apparatus - Dial	Accessories (if relevant) - Velcro cuffs: Adult size 54,5 x 14,5 cm Small adult size 42 x 13 cm, 1 tube, 2 tubes Obese size 70 x 15 cm, 1 tub Thigh size 70 x 22 cm, 1 tube	Velcro cuffFor better participation	No change
13	Ophthalmoscope	Detailed requirements - Magnification up to x15, in several selectable stages from direct vision to maximum magnification. Anti-reflection lens. At least 3 apertures and fixation star.	Magnification up to x15, in several selectable stages from direct vision to maximum magnification. Anti-reflection lens. apertures: slit, fixation star, large circle, medium circle, small circle, red-free filter, blue filter, polarization filter For better participation	No change
14	Ophthalmoscope	Lamp source - At least 2.5 V Xenon or LED light source	At least 2.5 V LED light sourceFor better participation	Lamp source - At least 2.5 V Xenon or 2.5 V LED light source
15	Ophthalmoscope	Aperture shape - Large spot, small spot, slit, central net, and red free	slit, fixation star, large circle, medium circle, small circle, For better participation	No change
16	Ophthalmoscope	Filters - Red-free, blue and polarization filters.	red-free filter, blue filter, polarization filterFor better participation	No change
17	Ophthalmoscope	Range of Lens -Range of lenses not smaller than -45D to +35D with steps not greater than 1D.	Range of lenses not smaller than plus (1-10, 12, 15, 20, 40); minus (1-10, 15, 20, 25, 30, 35)For better participation	Range of lenses not smaller than plus (1-10, 12, 15, 20, 40 or better); minus (1-10, 15, 20, 25, 30, 35 or better
18	Ophthalmoscope	Range of Lens - Should have rotating knob to control the intensity of the ophthalmoscope and should be used with filters that eliminate UV radiation (<400nm) and, whenever possible, filters that eliminate short- wavelength blue light (<420nm). Weight<4.5 kg required, fitting for mounting on wall, approx. size 450*350*100mm.	Should have rheotronic knob to control the intensity of the ophthalmoscope For better participation	Range of Lens - Should have rotating/rheotronic knob to control the intensity of the ophthalmoscope and should be used with filters that eliminate UV radiation (<400nm) and, whenever possible, filters that eliminate short- wavelength blue light (<420nm). Weight<4.5 kg required, fitting for mounting on wall, approx. size 450*350*100mm.
19	Ophthalmoscope	Accessories (if relevant) - Two spare bulbs.	Bulb not required When illumination is light source For better participation	No change
20	Laryngoscope with 4 blades - Adult	Spare parts (if relevant) - 2 sets of re-chargeable batteries. At least n. 1 spare light bulb compatible with the device provided.	2 sets of batteries. If battery is not rechargeable At least n. 1 spare light bulb compatible with the device provided. Kindly remove 1 Spare bulb because Light source is LED and LED is longlife bulb For better participation	Spare parts (if relevant) - 2 sets of re-chargeable batteries. At least 1 no spare light bulb compatible with the device provided if light source is not LED.
21	BP Apparatus- Automatic	Regulatory Approval / Certification - Should be BIS/ISO/FDA/ CE / UL approved product	European CE from Notified body For better participation	Regulatory Approval / Certification - Should be BIS/US-FDA/ European CE (from notified body) / UL approved product
22	BP Apparatus- Automatic	Detailed requirements - Cuff sizes: All in one size	Cuff sizes: Velro cuff Better	Detailed requirements - Cuff sizes: All in one size with velcro cuff

Sr.No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution/Amendment(s)
23	BP Apparatus- Automatic	Measurement Range - systolic (mm Hg), 60–250, 290 preferred for adults, 30–160 for children and 20– 120 for neonates. Diastolic (mm Hg), 30–180 adults, 10–150 paediatric, 10100 neonate. Mean arterial pressure (mm Hg), 30–250 adults, 30–160 children, 30–110 neonates. Pulse (beats per min), 30–150 adult and children, 30–180 neonates. Inflation pressure (mm Hg) 150–260 adults, 85–140 neonates; adjustable or automatically set preferred.	systolic (mm Hg), 60–255, Diastolic (mm Hg), 30–200 Pulse (beats per min), 40–200 Better	Measurement Range - systolic (mm Hg), 60–250, 290 preferred for adults, 30–160 for children and 20– 120 for neonates. Diastolic (mm Hg), 30–180 adults, 10–150 paediatric, 10100 neonate. Mean arterial pressure (mm Hg), 30–250 adults, 30–160 children, 30–110 neonates. Pulse (beats per min), 40–150 adult and children, 40–180 neonates. Inflation pressure (mm Hg) 150–260 adults, 85–140 neonates; adjustable or automatically set preferred. Above range is minimum approx. requirement, wider range than this is acceptable.
24	BP Apparatus- Automatic	Auto deflate pressure (mm Hg) - 300 adults, 150 neonates.	0-310mmHgBetter	No change
25	BP Apparatus- Automatic	Accuracy - Maximum error tolerance of +/- 2 mmHg	(+/- 3 mmHg)For better participation	Accuracy - Maximum error tolerance of +/- 3 mmHg
26	BP Apparatus- Automatic	Scope of supply - Includes Vinyl zippered bag Set with storage box, All-in one size cuffs, charger, battery set.	Includes Vinyl zippered bag Set with storage box, Velcro cuffs, charger, battery set.For better participation	Scope of supply - Includes Vinyl zippered bag Set with storage box, All-in one size velcro cuffs, charger, battery set.

Sr.No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution/Amendment(s)
27	<p>Allis Tissue Forceps- Large,Allis Tissue Forceps- Medium,Allis Tissue Forceps- Small,Curved Artery Forceps-Large,Curved Artery Forceps-Medium,Curved Artery Forceps-Small,Curved Scissors – Medium,Curved Scissors - Small,Lifting ForcepsMaggles Forceps Large,Maggles Forceps Medium,Maggles Forceps Small,B.P Knife Handle,Needle Holder- Large,Needle Holder-Medium,Needle Holder-Small,Plain Thumb Forceps- Large,Plain Thumb Forceps- Medium,Plain Thumb Forceps- Small,Sponge Holder,Stitch Cutting Scissors,Straight Artery Forceps- Large,Straight Artery Forceps- Medium,Straight Artery Forceps- Small,Straight Scissors- Large,Straight Scissors- Medium,Straight Scissors- Small,Toothed Thumb Forceps- Large,Toothed Thumb Forceps- Medium,Toothed Thumb Forceps- Small,Tuning Fork-Standard Frequency,Vaginal Retractor- Standard,Vaginal Speculum- Large,Vaginal Speculum- Medium,Vaginal Speculum- Small,Stapler Remover,Knife Dish with Lid -Medium, Knife Dish With Lid - Small, SS Bowl-Large, SS Bowl- Medium, SS Bowl-Small, SS Kidney Tray-Large, SS Kidney Tray-Medium, SS Kidney Tray-Small, SS Tray With Lid –Large, SS Tray With Lid - Medium, SS Tray With Lid -Small</p>	Sizes	As per the requirement of user Small, Medium and large size may varie or different, Hence we request you to kindly mention proper size in inches as per your requirement	Bidder should offer product as per standard Market practice. Product will be evaluated based on the sample submission.

Rest of tender terms and conditions remain same.

-sd-
(Head Procurement)