

ASSAM CANCER CARE FOUNDATION (ACCF)

3rd floor, V.K. Trade Centre, G.S. Road, Opp. Down-town Hospital,
Guwahati – 781022, Assam Ph: +91-90852 02020

E: procurement@accf.in | W: www.assamcancercarefoundation.org

**FOR SUPPLY, INSTALLATION & COMMISSIONING of “Medical Equipment, Surgical instrument and Hospital furniture” IN NEWLY
CONSTRUCTED CANCER CARE HOSPITALS AT DIFFERENT LOCATIONS IN ASSAM.**

Ref: ACCF/Medical Equipment/2021-22/36 Date 12.01.2022

CORRIGENDUM No. 1

Sr. No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution
1	Technical Specifications: Wheel Chair	Material; M.S Tubular framework fitted with S.S Seat and back.	M.S.TUBULAR FRAMEWORK FITTED WITH FOLDLE HEAVY DUTY REXINE COVER SEAT AND BACK.	<u>Wheel Chair Point No. 5</u> <u>Material:</u> Replaced as'M.S Tubular framework fitted with heavy duty rexine cover seat and back'.
2	Technical Specifications: Wheel Chair	Wheels: wo solid PU wheels and self-propelling S.S Loops. Two swivel castors 100 mm diameter in front.	TWO SWIVEL CASTOR WITH 150MM DIAMETER IN FRONT.	<u>Wheel Chair Point No. 6</u> <u>Wheels:</u> Solid PU wheels and self-popelling S.S Loops. Two swivel castors 150 mm diameter in front.
3	Technical Specifications: Wheel Chair	Arm rest; ABS plastic	PVC PLASTIC	<u>Wheel Chair Point No. 7 Arm Rest:</u> ABS/PVS Plastic

Sr. No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution
4	Technical Specifications: Examination couch	Storage cabinets; Three sliding door & three drawers for storage.	UPPER SECTION CONSISTING OF 3 DRAWERS FITTED WITH TELESCOPIC CHANNELS. LOWER SECTION WITH 3 CABINETS WHERE 2 OF THE CABINETS ARE FITTED WITH LOCKS.	<u>Examination couch Point No.2</u> <u>Storage cabinets:</u> Three sliding door & three drawers for storage. OR Upper section consisting of 3 drawers fitted with telescopic channels. lower section with 3 cabinets where 2 of the cabinets are fitted with locks.
5	Technical Specifications: Examination couch	Body Construction; body frame work made from 20G CRCA sheet & 20 mm x 40 mm x 16G.MS. rectangular Tubes. It should have fitted stainless steel legs & powder coated on all metal surfaces	BODY MADE FROM MACHINE PRESS 20G CRCA SHEET INBUILT SINGLE STEP STOOL FIITED IN BODY ALL DRAWERS AND CABINET MADE FROM CRCA SHEET 20G AND LEGS FITTED WITH ADJUSTABLE NYLON BUFFERS.	<u>Examination couch Point No.5</u> <u>Body Construction:</u> body frame work made from 20G CRCA sheet & 20 mm x 40 mm x 16G.MS. rectangular Tubes. It should have fitted stainless steel legs & powder coated on all metal surfaces. OR Body made from machine press 20G CRCA sheet inbuilt single step stool fiited in body all drawers and cabinet made from crca sheet 20G and legs fitted with adjustable nylon buffers.

Sr. No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution
6	Technical Specifications: Examination couch	Section dimension; Upper section of box approx. size 1220 mm L x 460 mm W x 630 mm H with three sliding drawers of approx. size 320 mm L x 430 mm W x 75 mm H. It should have lower section comprising of three cabinets of approx. inside size 350 mm L x 440 mm W x 430 mm H with separate doors.	UPPER SECTION TWO DRAWER 330MM W X 160MM HT CENTER DRAWER 510 MM W X 160 MM HT LOWER SECTION TWO CABINET WITH LOCK 355 MM HT X 330 MM W AND CENTER DRAWER 550MM W X 275MM HT	<u>Examination couch Point No.7</u> <u>Section dimension:</u> Upper section of box approx. size 1220 mm L x 460 mm W x 630 mm H with three sliding drawers of approx. size 320 mm L x 430 mm W x 75 mm H. It should have lower section comprising of three cabinets of approx. inside size 350 mm L x 440 mm W x 430 mm H with separate doors. OR Upper section two drawer 330MM w x 160MM ht center drawer 510 MM w x 160 MM ht lower section two cabinet with lock 355 MM ht x 330 MM w and center drawer 550MM w x 275 MM ht
7	Technical Specifications: Height adjustable couch with side rails and wheels for USG room / Echo room: -	Body Construction; body frame work made from 20G CRCA sheet & 20 mm x 40 mm x 16G.MS.rectangular Tubes. It should have fitted stainless steel legs & powder coated on all metal surfaces	YES LEGS FITTED WITH ADJUSTABLE BUFFERS.	Group 'A', Item No. 4- Height Adjustable Couch: This item is dropped from the tender. Pls put zero (0) in the financial bid as we are not revising the financial bid format.
8	Technical Specifications: Height adjustable couch with side rails and wheels for USG room / Echo room: -	Foot steps; Sliding foot step under the front side of lower middle cabinet	SINGLE STEP STOOL SEPARATLY GIVEN	This item is dropped from the tender. Pls put zero (0) in the financial bid as we are not revising the financial bid format.
9	Group 'A', Item No. 4- Height adjustable couch with side rails and wheels for USG room / Echo Room	Wheels and Side railings; Required with locking mechanism	WHEELS & SIDE RAIL ARE NOT AVAILABLE ON THIS MODEL. KINDLY DISCUSS WITH THE USER DEPARTMENT FOR CLARIFICATION.	This item is dropped from the tender. Pls put zero (0) in the financial bid as we are not revising the financial bid format.

Sr. No.	Clause No/ Query Ref	Clause as per RFP	Query	Query Resolution
10	Average Annual Turnover	Clause No.5.1.3 The Bidder should have an average annual turnover of Rs. 3.00 Cr or more in the last three (3) financial years (i.e., 2017-18, 2018-19, 2019-20 or 2018-19, 2019-20, 2020-21)duly certified by the Chartered Accountant as per the format at Format T8.	For the above 2 mentioned groups B & D the qualifying criteria for the tender is very high, as items mentioned in the tender tend to be of a relatively low value. So we would request you to kindly revise the average overall Turnover criteria to 1 crore in the last 3 financial years.	Only for GROUP B & D-Turnover of 1 crore or more in the last 3 financial years.i.e., (2017-18, 2018-19, 2019-20 or 2018-19, 2019-20, 2020-21). However for rest of the Groups average annual turnover of 3 CR or More is required.
11	Average Annual Turnover	Clause No.5.1.3 The Bidder should have an average annual turnover of Rs. 3.00 Cr or more in the last three (3) financial years (i.e., 2017-18, 2018-19, 2019-20 or 2018-19, 2019-20, 2020-21)duly certified by the Chartered Accountant as per the format at Format T8.	For the above 2 mentioned groups B & D the qualifying criteria for the tender is very high, as items mentioned in the tender tend to be of a relatively low value. So we would request you to kindly revise the average overall Turnover criteria to 1 crore in the last 3 financial years.	Only for GROUP B & D-Turnover of 1 crore or more in the last 3 financial years.i.e., (2017-18, 2018-19, 2019-20 or 2018-19, 2019-20, 2020-21). However for rest of the Groups average annual turnover of 3 CR or More is required.

Rest of tender terms and conditions remain same.

-SD-

(Head Procurement, ACCF)