

ASSAM CANCER CARE FOUNDATION (ACCF)
3rd floor, V.K. Trade Centre, G.S. Road, Opp. Down Town Hospital,
Guwahati – 781022, Assam Ph: +91-90852 02020
Email: procurement@accf.in | W: www.assamcancercarefoundation.org

Date: 16.4.2021

ONLINE TENDER FOR SUPPLY INSTALLATION AND COMMISSIONING OF NURSE CALL SYSTEM

Ref. Tender No: ACCF/NCS/20-21/14 Dated 11.3.2021

CORRIGENDUM No. 3

Amendment & clarifications for the above said bid document are being issued as follows along with extension of bid submission date. Now, bid submission date is 26.4.2021. Bidders are requested to take note of following clarification/amendments while quoting the bid.

Sl no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
1	Page 6 Table (2)	Within 30 days from the date of issue of Work Order/Supply Order.	L-1 = 120 days L-2= 90 days L-3= 60 days (subject to site is ready for installation)	After receipt of order, it takes some time to mobile resources for installation and conducting and wiring takes sometimes.	Supply, installation and commissioning should be as follows from the date of site handover: A. L1 site: 60 days B. L2 sites: 50 days C. L3 sites: 35 days
2	4.2.1	Technical specifications: <i>Completion of Installation and Commissioning.</i> 45 days from date of issuance of Purchase/Work Order.	L-1 = 120 days L-2= 90 days L-3= 60 days (subject to site is ready for installation)	Completion time should be based on no of Beds as volume of depends on it.	Wherever, in bid document, delivery, installation, and commissioning is mentioned, above shall prevail.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
3	5.1.1(ii) (iii)	<p>Import License (In case of Importer only).</p> <p>In case of Importer or authorised dealer, they will have to furnish the manufacturer's authorization form from the original equipment manufacturer (OEM) as per Format T7</p>	Delete entire clause	Nurse Call Systems are being designed and manufactured in India and hence no need for import them and west foreign exchange. Availability of spares, after sale services and its cost may cause long term problems.	It is clarified that all indigenous or foreign make manufacturers/their authorized dealers can participate in the tender.
4	5.1.2	Authorized Distributors/Dealer are also eligible to participate in the bid provided:	Delete entire clause	Dealers or traders will not be able to provide technical support, modifications, or service in long term.	No change
5	6.1.2	Modification of contract: If necessary, the Tender Inviting Entity may, by a written order given to the successful bidder at any time during the currency of the contract, amend the contract by <u>making alterations and modifications within the general scope of</u>	Delete entire clause	After finalization of contract, amendments/ modifications to the scope of contract, its specifications, and other terms can be treated as extended advantages to contract awardee and deprivation to	It is clarified and amended that only minor changes might be made in the contract with mutual consent without materially affecting tender conditions.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
		<p><u>contract, in any, one or more of the followings:</u></p> <p>a) Specifications, drawings, designs, etc., of the Nurse Call System to be commissioned at respective health facility/hospital,</p> <p>b) Mode of Demonstration/Quality Inspection</p> <p>c) Incidental services to be provided by the successful bidder</p> <p>d) Mode of Installation</p> <p>e) Any other term(s) of the contract, as felt necessary by the Tender Inviting Entity depending on the merits of the case</p>		<p>other prospective bidder.</p> <p>It may effect transparency in the tender process.</p>	
6	6.3.6) Installation & Commissioning :	The electrical power supply point will be provided by the purchaser at the room where the equipment will be installed but the wiring and electrical fittings inside the room and accessories if any required for installation and	The electrical power supply point will be provided by the purchaser at the room where the equipment	Please note that during the conduit fixing and Cat6 cable lying part of NCMS, Chase Cutting of wall & filling should be provided by the civil team. So we are requesting you to	No change. Chase cutting of wall and filling to bring back to its original position is also the responsibility of the bidder.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
-------	----------------	--------------	------------------------------------	---------------	-------------------------------

		commissioning of the equipment from the power supply point to the point of actual installation or any other civil work required for installation of the System will be provided by the supplier without any extra cost. The contract price as offered in the price bid and agreed shall be all inclusive. No separate payment shall be made other than the contracted price.	will be installed but the wiring and electrical fittings inside the room and accessories if any required for installation and commissioning of the equipment from the power supply point to the point of actual installation or any other civil work required for installation of the System it will be provided by the ACCF. The contract price as	please clarify this point.	
--	--	--	---	----------------------------	--

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
			offered in the price bid and agreed shall be all inclusive. No separate payment shall be made other than the contracted price.		
7	6.4) Payment 6.4.2	90% of the cost of the equipment installed (excluding CMC Cost, if any) + 100% tax shall be paid to the supplier on receipt (stock entry), installation and demonstration of the item.	70% of the cost of the equipment supplied(excluding CMC Cost, if any) + 100% tax shall be paid to the supplier on receipt (stock entry),	The job is in two part. 1 is supply of equipment and 2 nd is it's installation and commissioning. We request you for amendments to justify for both supplier and ACCF. This is standard for such installation.	Sub-clause 6.4.2 of Clause 6.4: Payment is amended as follows:- 80% of the cost of the equipment (excluding CMC Cost, if any) + 100% tax shall be paid to the supplier on receipt of material (Stock entry) subject to submission of equivalent amount of PBG. 10% of the payment of equipment will be made after installation of equipment. Balance 10% after receipt of certificate on working status of the equipment
8	6.4.3	The balance 10% of the payment of equipment will be made after receipt of certificate on working status of the equipment	The balance 30% of the payment of equipment will be made		

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
		from the consignee after 8 weeks of installation and commissioning of the equipment.	after receipt of certificate on working status of the equipment from the consignee after 2 weeks of installation and commissioning of the equipment.		from the consignee after 8 weeks of installation and commissioning of the equipment. Note: If supplier fails to submit PBG, as mentioned above, then 90% payment shall be made after installation, commissioning and demonstration of the system.
9	6.10) Imported Equipment	6.10.1 ACCF shall no way involve in the import of the equipment from foreign countries, if such equipment is manufactured outside the country. It shall be the sole responsibility of the bidder to import the equipment offered by paying the requisite consideration in foreign currency and following the stipulations issued by the Government of India, from time to time, in the import	Delete entire clause	When manufacturers of such items are Available in India with proven record Import of such items may be considered as against our aim of Make in India. Eligibility criteria must be restricted to Indian Manufacturers only.	No change.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
	2.6.1.2 Page No 5, Point No 5	<p>of equipment, especially when the import is from hostile nations.</p> <p>The bidder can furnish the EMD (i.e., Bid Security) amount in either of the form given below: a) DD or Fixed Deposit Receipt (FDR) duly lien marked in favour of ACCF or b) Irrevocable Bank Guarantee (BG) issued in favour of ACCF by any scheduled bank in India as per the format given in Annexure-V. c) Online payment (NEFT/RTGS/IMPS) d) Submission of key-documents in originals on 02.04.2021</p>	<p>Add: d) Manufacture registered with NSIC for Nurse Call Monitor System may be exempted for summation of EMD on providing valid copy of NSIC registration certificate.</p> <p>Submission of key-documents in originals on 12.04.2021</p>	<p>As per MSME policy of Govt. of India, NSIC units be exempted from EMD amount</p> <p>Due to festivals, year ending issue, State Elections, banking services will remain effected.</p>	

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
10	EMD 2.6.1.2 Page No 5, Point No 5	The bidder can furnish the EMD (i.e., Bid Security) amount in either of the form given below: a) DD or Fixed Deposit Receipt (FDR) duly lien marked in favour of ACCF or b) Irrevocable Bank Guarantee (BG) issued in favour of ACCF by any scheduled bank in India as per the format given in Annexure-V . c) Online payment (NEFT/RTGS/IMPS) d) Submission of key-documents in originals on 02.04.2021	Add: d) Manufacture registered with NSIC for Nurse Call Monitor System may be exempted for summation of EMD on providing valid copy of NSIC registration certificate. Submission of key-documents in originals on 12.04.2021	As per MSME policy of Govt. of India, NSIC units be exempted from EMD amount Due to festivals, year ending issue, State Elections, banking services will remain effected.	No change. Clear 7 days shall be given from the date of issue of amendments for bid submission.
11	6.14.3	Venue of Arbitration: The venue of arbitration shall be the place from where the contract has been issued, i.e., Mumbai, Maharashtra.	Guwahati/ Assam	When the work is to be done in state of Assam, having jurisdiction of Mumbai / High court of Maharashtra may	Clause 6.14.3 amended as: Venue of Arbitration: The venue of arbitration shall be the place from where the contract has been issued, i.e., Guwahati, Assam.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
				be replaced by Guwahati / Assam.	Clause 6.15.2: Place of jurisdiction shall be Guwahati (Assam).
12	6.15.2	All disputes arising out of this bid will be subject to the jurisdiction of courts of law in Mumbai / High court of Maharashtra.	Guwahati/ Assam		
13	7.1 Technical Specifications	G) Enhanced Device Intelligence a) Communications and patient terminals should be equipped for receiving up to 32 audio streams (e.g. radio programmes or in-house programmes). Furthermore, every system switch should contain an interface to an external TV device, which can be assigned to a specific room via the configuration.	Delete entire clause	This is against the basic objective of Program "create patient-centric cancer institutions to deliver standardized and affordable care closer to patients " homes (hereinafter referred to as "Program").	G) Enhanced Device Intelligence: This requirement is deleted.
14	7.2.1) Quality Standard	For Manufacturer: The manufacturer shall have following ISO Certification a) ISO 14001-2004 EMS (Environmental Management System) b) ISO 18001-18001-2007 (Occupational Health & Safety)	For Manufacturer: The manufacturer shall have following ISO Certification 9001:2015.	Manufacturing units current applicable ISO is 9001:2015 ISO 14001-2004 EMS (Environmental a) Management System)- is not applicable for	Clause 7.2.1-Quality Standard Replaced as: For Manufacturer: The manufacturer shall have following ISO Certification a) ISO 9001-2000 (QMS) b) International standard IEC 60601-1 or UL2601-1 or CAN/CSA-C22.2 No. 601.1-

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
		c)ISO 9001-2000 (QMS)	Delete other points	electronics and software manufacturing units. b)ISO 18001-18001-2007 (Occupational Health & Safety) is not applicable for electronics and software manufacturing units	M90 and c) ISO 13485: 2016 or 21CFR820
15	7.1.2. (B) System Architecture	<p>System Architecture Superordinate and/or centralised controller devices are not permitted for safety reasons.</p> <p>Preferred Make mentioned are :-Schrack, Ackerman, Austco, Ascom,Tunstal</p> <p>These matrix's are also not fully compiling to technical specifications as published and still many objections are raised by many intending participants</p>	<p>Please note that all the brands are manufactured outside India. This goes against our National Policy for Make in India.</p> <p>We appeal for immediate withdrawal of such matrix till tender is not finalized and contract is awarded.</p>		All companies, who meets the technical specifications and bid conditions are accepted irrespective of brands/photographs mentioned in the bid document. Photographs are only for illustration purpose.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
		On 19 th March, we observed that BOQ-Nurse Call Assam L-1,L2,L3 matrix is published. These matrix's are published with photos of a particular maker. This creates that authorities have already selected particular make and supplier.	We strongly object to this and in our opinion, this is not fair with other intending participants.		
16	7.1.2. G)Enhanced Device Intelligence	a) Communications and patient terminals should be equipped for receiving up to 32 audio streams (e.g. radio programmes or in-house programmes). Furthermore, every system switch should contain an interface to an external TV device, which can be assigned to a specific room via the configuration.	Delete entire clause	This is against the basic objective of Program "create patient-centric cancer institutions to deliver standardized and <u>affordable care closer to patients</u> " homes (hereinafter referred to as "Program").	Enhance Device Intelligence requirement is deleted.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
17	7.1.2.C) Patient Handset with Speech	Patient Handset with speech for communications, consist of: <ul style="list-style-type: none"> • Splash resistant case • Call button with nurse symbol with integrated finder and reassurance light • Loudspeaker, microphone, headphones socket, • Integrated contactfree smart card reader, • Mechanism for inserting a smart card, • Controlling 2 independent lighting circuits, • Plastic case and membrane keypad in antimicrobial material 	Delete entire clause	This item is not mentioned in given BOQ for nurse call system. It may be deleted.	Amendment: Patient handset without speech is required. Speech facility not required.
18	7.1.6 B) Connection module for Nurse Station Terminal & Patient Handset with Speech	For installation in a double switchbox or in a media duct, both for nurse station terminals; consisting of: <ul style="list-style-type: none"> • Mounting rail, • Connection circuit board with an RJ45 socket for the uplink from the system switch, • RJ45 socket marked in colour and with measures to ensure that the patient terminal is correctly connected, including the 	Delete entire clause	This item is not mentioned in given BOQ for nurse call system. It may be deleted.	Clause 7.1.6 B)- Entire clause is deleted.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
		<p>mechanical parts required for the auto disconnect mechanism,</p> <ul style="list-style-type: none"> •RJ45 for connecting a laptop belonging to the patient: The patient thereby has access to the WWW via the system's Internet server. •DIN socket for connecting a diagnostic device •Fastening frame made of plastic for attaching the connection module without needing screws 			
19	7.1.6 I) Room terminal –For Single Bed Rooms	<p>For installation in all important rooms in the ward in which the possibility to communicate is required or desired, with integrated functional components and operating membrane keypad with coloured fields, LC display and symbols; consisting of:</p> <ul style="list-style-type: none"> •Membrane keypad for operation, comprising of: •Call button (red with nurse symbol) with integrated finder and reassurance light 	Delete entire clause	This item is not mentioned in given BOQ for nurse call system. It may be deleted.	NO CHANGE. To be quoted.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
		<ul style="list-style-type: none"> •Doctor call button (blue with doctor call symbol) with integrated reassurance light, •Presence button (green) with control LED, •Presence button (blue for the doctor) with control LED, •LC Display •Room terminal shall display the nurse call from other rooms/ patients, in case nurse acknowledge the call from room terminal. 			
20	7.1.6 J) Doctor Call & Cancel Button with sounder -for Wards	It shall consist of: <ul style="list-style-type: none"> •1 doctor call button (Blue) with a finder light •1 cancel button •Integrated piezo buzzer 	Delete entire clause	This item is not mentioned in given BOQ for nurse call system. It may be deleted.	NO CHANGE. To be Quoted
21	2.8.1	Validity of bid must be 180 days	Validity of bud must be 90 days from date of last submission	Considering the fluctuation in euro since our product is imported	No Change
22	Section 3 tender details	Performance security 10% of contract value	We request for the same to be 2.5% of the order	The same needs to be valid for 2 years from handover for	Performance Security: amended as 5% of contract value (without GST). At the time of contract agreement,

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
			value from the date of handover of each site	each site which is quite a long period	selected bidder need to submit 2% of performance security and balance 3% of order value immediately (within 15 days) at the time of the placement of site wise Purchase Order(s).
23	4.2.1	Completion of Installation and	Guwahati - 170 beds (4 months)	considering cabling and conduiting to be included in our scope of works it will be a challenge to finish all site in 45 days. Also for sites above 75 beds based on our past experience it always is a challenge to finish the same in one visit considering dependency of our works on many agencies. Also the months mentioned in desired amendment column is	Delivery, installation and commissioning time shall be as explained above.
		Commissioning.- 45 days	Dibrugarh - 211 beds (4.5 months)		
			Barpeta - 235 beds (5 months)		
			silchar - 235 beds (5 months)		
			Diphu - 235 beds (5 months)		
			Jorhat - 61 beds (2 months)		
			Tezpur - 61 beds (2 months)		

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
-------	----------------	--------------	------------------------------------	---------------	-------------------------------

			lakhimpur - 61 beds (2 months)	minimum time period we expect.	On expiry of Warranty , ACCF will have the right to enter into an AMC/ CMC with the vendor . Vendor/ Bidder to provide yearly AMC & CAMC values valid for maintaining the equipment for a period of 5 Years post 2 years' warranty period.
			mangaldoi - 61 beds (2 months)		
			kokrajhar- 61 beds (2 months)		
24	4.2.3.	CMC/AMC period (wherever applicable)	5 years CMC after warranty . Couldn't understand the point in the tender please clarify		
25	6.4	No advance payments towards cost of item supplied and installed will be made to the Supplier.	We request for 50% advance payment	Ours is an imported product and our principle terms are 100% advance	Payment terms as modified above.
		90% of the cost of the equipment installed (excluding CMC Cost, if any) + 100% tax shall be paid to the supplier on receipt (stock entry), installation	we request for 40% payment against delivery		

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
		and demonstration of the item.			
26		The balance 10% of the payment of equipment will be made after receipt of certificate on working status of the equipment from the consignee after 8 weeks of installation and commissioning of the equipment.	We request for release of 7.5% against handover as we have requested for 2.5% as performance security	Performance security is 10% as per tender so even after handover our payment will be withheld for the term of performance security	5% performance security as amended above.
27	6.7.4	- AMC will be 2.5% and CMC would be 5% of purchase value (excluding GST) of the equipment. Thereafter, 5% increase every year over previous year value. Applicable GST shall be paid extra.	We request you to consider AMC Last 6% of order value and CMC as 9% of order value	Considering manpower overheads post 3 years from now and also the general % of AMC and CAMC in the market we are suggesting these %	No change
28	6.17.5	Liquidated Damages:- If the successful bidder fails to install the system within the time frame(s) prescribed in the contract, ACCF shall, without prejudice to other rights and remedies	We request to not keep any LD clause in the tender	Reason being is our works are dependent on many agencies like BHP, UPS, civil works, painting,	Clause 6.17.5 amended as: Liquidated Damages:- If the successful bidder fails to install the system within the time frame(s) prescribed in the contract, ACCF shall, without

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
-------	----------------	--------------	------------------------------------	---------------	-------------------------------

		<p>available to it under the contract, deduct from the purchase order price as liquidated damages, a sum equivalent to 2% of the order value of the item to be supplied and (or) installed, per each week of delay or part thereof until actual commissioning or performance subject to a maximum of 6%. ACCF reserves the right to allow an additional penal period of 2 (two) weeks beyond the normal penal period (3 weeks) on the written request of the supplier with the condition that liquidated damage @ 2.5% will be charged for each week or part thereof during the extended penal period.</p>		nurse station readiness etc.	<p>prejudice to other rights and remedies available to it under the contract, deduct from the purchase order price as liquidated damages, a sum equivalent to 1% of the order value of the item to be supplied and (or) installed, for each week of delay or part thereof until actual commissioning or performance subject to a maximum of 5%. ACCF reserves the right to allow an additional penal period of 2 (two) weeks beyond the normal penal period (5 weeks) on the written request of the supplier with the condition that liquidated damage @ 2% will be charged for each week or part thereof during the extended penal period.</p>
--	--	--	--	------------------------------	---

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
29	7.1.6(B)	(B) Connection module for Nurse Station Terminal & Patient Handset with Speech	No amendment needed	Any IP based system whether its CCTV, FIRE, Nurse call, backbone switches etc. if has RJ45 connector then system will be IP based and also helps in remote access of the system	7.1.6 (B) amended as: Connection module for Nurse Station Terminal & Patient Handset without Speech
		· Connection circuit board with an RJ45 socket for the uplink from the system switch,			For installation in a double switchbox or in a media duct, both for connecting for connecting nurse station terminals; consisting of:
		· RJ45 socket marked in colour and with measures to ensure that the patient			<ul style="list-style-type: none"> • Mounting Rail • Connection circuit board with an RJ45 socket for the uplink from the system switch, • RJ45 or equivalent socket marked in colour and with measures to ensure that the patient Hand switch is correctly connected, including the mechanical parts required for the disconnect mechanism,

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
					<ul style="list-style-type: none"> Deleted Deleted Deleted
30	7.1.6.(H)	conforming to VDE0834 and suitable for fitting on to an installation case, consisting of:	No amendment needed	since the BoQ has requirements of different calls like nurse call, nurse presence, toilet call, nurse assistance call and code blue call we have kept lamp with 5 different indications	Amendment: 4/5 light chambers are accepted.
31		<ul style="list-style-type: none"> 5 light chambers with light reflectors for homogeneous illumination 	Let the same be 5 lights as we don't have 4 light chamber	5 lights are for code blue, nurse presence, second nurse call, toilet call, nurse call	Amendment: 4/5 light chambers are accepted.
32	2.24	The contract will be awarded to the lowest evaluated responsive bidder (the one who is lowest in total BoQ quote), adjudged vide the financial bid evaluation of all the technically qualified bidders provided:	L &T must be considered as just another competition bidder	L &T must be considered as just another competition bidder	No change

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
		a) L&T Limited (the principal construction contractor of the project) refuses to execute tendered item on its own at L1 price and			
33	Installation certificate, 2 months performance certificate and performance certificate		Can we submit the same as per letters received from our client	We need some extension in tender submission date in case the same are needed to be submitted as per tender format considering availability of our clients as it needs seal from our hospital	Installation and Performance certificates to be submitted by selected bidder after installation and commissioning.
34	Not mentioned in tender	chase cutting and civil work in whose scope	the same must be in scope of civil contractor	They have agencies who can do the cementing and color work	Chase cutting and related civil work shall be carried out by the selected bidder.
		Installation of wall box for toilet and lamp modules in whose scope			Installation of wall box for toilet and lamp modules is in the scope of Nurse Call Vendor.
35	Not mentioned in tender	Bed head panel cut out and UPS	Please inform hospital	Cut out in scope of BHP vendor and UPS needed on	Bed Head Panel shall be supplied by MGPS vendor. NCS vendor shall give its cut-

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
			wise where all and which areas bed head panel are to be considered and cut outs shall be in whose scope	each nurse station near controller	out size to MGPS vendor to make a suitable cut to fix NCS device/connection. UPS is in the scope of ACCF.
36			If BHP is not to be considered in those areas will the installation be done in wall ?	Installation of BHP in modules is better considering maintenance perspective and aesthetics	Wherever BHP is not installed, then installation shall be done on wall.
37	Not mentioned in tender	Return of EMD	On awarding the tender		EMD shall be returned once selected bidder submits the performance security. Technically disqualified bidders shall get their EMD back after opening of Financial bids.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
38	BOQ clause 6	In your tender documents the Technical specifications given are of VDE-0843 standards only and on some places only UL is mentioned	We request kindly amend your Tender documents and it should be mentioned that the Bidder should follow VDE-0843 or UL-1069 standards (Where ever applicable) for supply of Nurse Call system.	In Nurse Call system leading companies follow mostly two standards world - wide, one is VDE-0843 and the second one is UL-1069.	Either VDE-0843 or UL-1069 is accepted. However, only one single standard (either VDE-0843 or UL-1069) is allowed to quote i.e. no mix and match.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
39	BOQ clause no D	In specifications of Toilet Units VDE 0834 certified/ approved Pull Cord with cancel Button - For WC	Our submission is that the toilet units be VDE 083/UL 1069 certified /approved pull cord with cancel button and should be IP68 rating for water and dust proof capability	Because there is a chance that call cord may be damaged because of water and for patient safety.	Toilet units be VDE 083/UL 1069 certified /approved Pull cord with cancel button should be IP68 rating for water and dust proof capability.
40	All items in BOQ	List of makes given are Make:- Schrack, Ackerman, Austco, Ascom,Tunstall or equivalent	We request for inclusion of name of our Principal Company Amtek Rauland, USA to be included in the list of makes given in the BOQ sheet item wise and other places of the tender documents		Equivalent make shall be considered who meets technical specs.

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
41	BOQ clause G	Supply of Backbone Switch Floor Switch: Uplink connections: 2 x 1Gb uplink ports, RJ45 (GBIC capable, possibility for LWL) Connections downlink: 24 x 100Mb downlink ports, RJ45 Switch must have all the Layer 2 and Layer 3 function which is mentioned in the tender Document	Our submission is that it is not required in our system, again this component is specific for a particular make. So it should be optional.	Specific to particular make.	BoQ Point No. G: This is a third-party item. Equivalent L2,L3 switch shall be considered.
42	7.1.2, page 52	Deviations from specifications.	Our product will not match the exact specifications of the tendered NCS our make is "Process care systems"	According to make in India Policy , priority to be given to Indian manufacturers	Already explained above.
43	6.4 , page 38	Can we submit running bills for material?	we would like to supply material after site is ready in phased manner.	If we supply material and site is not ready, the entire inventory will be blocked and we won't be able to hand over the site	Payments terms shall be as explained above. No running bills shall be accepted. This is a turnkey project. However, in case work is partially completed and vendor can not install the remaining system attributable to ACCF (i.e. due to site conditions),

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
					then running bills can be accepted for payment.
44	5.1.1, page 33	We can submit PO copies for proof of supply.	Please accept PO copies as proof of supply wherever work completion letter is not submitted.	we do not have work completion letters from every customer.	PO copies with Invoice raised can be submitted along with proof of payment. Generally, all private hospitals issue work completion certificates. In case of Govt. supply, pls submit Work Order copies along with payment proof.
45	page 5	Due date extension	Kindly provide due date extension by 7 days. i.e 9-4-2021	We need time to collect documents.	Already extended.
46	2.6.1.2, page 6 and 15	EMD exemption for MSMEs	For MSMEs, EMD to be waived off. Bid declaration can be given instead.	Govt. of India policy	No change
47	Format T5 declaration. page no.69	Stamp paper value is not mentioned	Stamp paper value to be mentioned	NIL	Rs. 100/- non-judicial stamp paper.
48	Annexure-2: Financial Bid Summery	In the given BoQ: site wise qty of GI/PVC Conduit and Cat 6 cabling is not given, hence not able to give site wise price quote in Financial Bid summery.	Pls give site wise qty of GI/PVC conduit and Cat 6 cabling	For site wise financial quote, site wise quantity of CG/PVC conduit and Cat 6 cabling is	Exact site wise quantity is not known. For calculation purpose, take 125 meter/site for GI/PVC conduit and 500 mtr/site for Cat 6 cabling. These nos. are just for

SI no	BID clause no.	Query raised	Desire amendment and clarification	Justification	ACCF Clarification/Amendments
-------	----------------	--------------	------------------------------------	---------------	-------------------------------

				required similar to other line items.	calculation/comparison purpose.
49	Notice Inviting Tender: Calendar of Events	<p>Existing Clause:</p> <ul style="list-style-type: none"> - Last date and time of bid submission (Online): 16th April 2021 till 1700 Hrs - Last date and time of submission of physical Key documents (Financial & Technical Proposal online). Financial to be submitted online only. Key documents required in Hard copies in Technical bid: 16th April 2021 at 17:00 Hrs - Bid Opening Date (online & offline opening): 16th April 2021 at 17:30 Hrs 	Extension of bid submission date is required		<p>Amended dates:</p> <ul style="list-style-type: none"> - Last date and time of bid submission (Online): 26th April 2021 till 1700 Hrs - Last date and time of submission of physical Key documents (Financial & Technical Proposal online). Financial to be submitted online only. Key documents required in Hard copies in Technical bid: 26th April 2021 at 17:00 Hrs - Bid Opening Date (online & offline opening): 26th April 2021 at 17:30 Hrs

Rest of tender terms and conditions remain same.

--sd--
(Head Procurement)