

ASSAM CANCER CARE FOUNDATION (ACCF)  
 3rd floor, V.K. Trade Centre, G.S. Road, Opp. Down Town Hospital,  
 Guwahati – 781022, Assam Ph: +91-90852 02020  
 E: info@accf.in | W: [www.assamcancercarefoundation.org](http://www.assamcancercarefoundation.org)

## **ONLINE TENDER FOR ENTERING INTO RATE CONTRACT FOR SUPPLY OF MEDICAL EQUIPMENT**

**Ref: Tender No: ACCF/MEDEQUIPMENTS/19-20/01**

**Dated: April 7, 2020**

### **COORIGENDUM No. 3**

Amendments, as follows are being made to above mentioned tender document:

Sr. No.	Tender Clause No.	Description					Amendment		
1	1.2.8	Item No.	Name of Item	Approx. Qty* for Diphu and Dibrugarh Centers	Approx. Qty for Silchar Day care Centre	EMD Amount	Name of the item	Revised Qty	Revised EMD
		1	Infusion pumps	15	1	13000	Infusion pumps	6	3600
		2	Syringe pumps	15	8	14000	Syringe pumps	18	9000
		3	Patient monitor	44	3	61000	Patient monitor	13	20800
		4	ECG	2	1	7000	ECG	6	14400
		5	Defibrillator	2	1	15000	Defibrillator	5	21000
		6	Portable ventilator	2	0	20000	Portable ventilator	2	24000
							Pulse Oxymeter	5	3000
							Portable suction machine - high vac	10	4000
		7	Pulse Oxymeter	4	0	1500	Vertical Autoclave	5	11000

		<table><tr><td>8</td><td>Portable suction machine - high vac</td><td>3</td><td>1</td><td>2500</td></tr><tr><td>9</td><td>Vertical Autoclave</td><td>2</td><td>1</td><td>100000</td></tr><tr><td>10</td><td>Oxygen cylinders with flow meters</td><td>8</td><td>3</td><td>45000</td></tr><tr><td>11</td><td>Biosafety Cabinet</td><td>2</td><td>1</td><td>20000</td></tr></table> <p>At present, these quantities are for ACCF Cancer Care Centres at Dibrugarh, and Diphu which are required to be supplied within two weeks from the date of placement of purchase order. For other center(s), supply period is 6 weeks from the date of placement of purchase order(s), if any. Bidders are required to quote for all quantities for an item (i.e. not center wise).</p>	8	Portable suction machine - high vac	3	1	2500	9	Vertical Autoclave	2	1	100000	10	Oxygen cylinders with flow meters	8	3	45000	11	Biosafety Cabinet	2	1	20000	<table><tr><td>Oxygen cylinders with flow meters</td><td>13</td><td>2600</td></tr><tr><td>Biosafety Cabinet</td><td>4</td><td>28000</td></tr></table> <p>Supply period is 40 days from the date of placement of purchase order(s), if any. Bidders are required to quote for all quantities for an item.</p>	Oxygen cylinders with flow meters	13	2600	Biosafety Cabinet	4	28000
8	Portable suction machine - high vac	3	1	2500																									
9	Vertical Autoclave	2	1	100000																									
10	Oxygen cylinders with flow meters	8	3	45000																									
11	Biosafety Cabinet	2	1	20000																									
Oxygen cylinders with flow meters	13	2600																											
Biosafety Cabinet	4	28000																											
	1.3.1	1.3.1 The contract with the successful bidder shall be for a period of <i>Six (6) months</i> from the date of signing of the contract. The management of ACCF may, if feel necessary, extend/renew the contract for a maximum period of three (3) months on similar terms and conditions and with mutual consent. However, ACCF reserves the right to terminate the bid process or the contract at any time without assigning any reason thereof.	1.3.1 The contract with the successful bidder shall be for a period of Twelve (12) months from the date of signing of the contract. The management of ACCF may, if feel necessary, extend/renew the contract for a maximum period of three (3) months on similar terms and conditions and with mutual consent. However, ACCF reserves the right to terminate the bid process or the contract at any time without assigning any reason thereof.																										

	1.12.1	<p><b>1.4 SECURITY DEPOSIT</b></p> <p>1.12.1 The successful bidder, within 10 days of signing of the agreement for Diphu and Dibrugarh quantities and thereafter upon receiving the every purchase order, shall be required to submit Performance Security Deposit of 5% of the order value in the form of bank guarantee from any Indian scheduled bank in favour of "Assam Cancer Care Foundation" valid at least for 60 days beyond the warranty period. However, if the supplier fails to execute the order or fails to perform the obligations under the contract or the purchase order, in addition to other penal actions, the bank guarantee shall be encashed and the amount will be forfeited.</p>	<p><b>1.4 SECURITY DEPOSIT</b></p> <p>1.12.1 The successful bidder, upon acceptance of rates, within 10 days, shall submit 2% of rate contract value as Security Deposit of the approved quantities valid for the entire duration + 60 days and shall also sign the rate contract agreement for the approved tendered quantities. Thereafter upon receiving every purchase order, shall be required to submit Performance Security Deposit of 3% of the order value in the form of DD/Bank guarantee from any Indian scheduled bank in favour of "Assam Cancer Care Foundation" valid for at least for 60 days beyond the warranty period. If supplier fails to honor the rate contract, 2% Security Deposit shall be forfeited apart from other penal actions. Further, if the supplier fails to execute the order or fails to perform the obligations under the contract or the purchase order, in addition to other penal actions, the total equivalent 5% Security Deposit against that particular equipment or 2% total security deposit (which was submitted along with the RC Agreement) (whichever is lower) shall be encashed and the amount will be forfeited. However, ACCF reserves the right to forfeit complete 2% Security Deposit or 5% Security Deposit as the case may be.</p>
--	--------	--	--

	1.13.5	Firm Delivery Schedule: First purchase order shall be placed immediately for Dibrugarh and Diphu and delivery of these items should be made within 2 weeks from the date of purchase order. For rest of purchase order(s), delivery shall be made within 6 weeks from the date of purchase order(s). Cross conditions such as "SUBJECT TO AVAILABILITY" "SUPPLIES WILL BE MADE AS AND WHEN SUPPLIES ARE RECEIVED" etc., will not be considered under any circumstances and the tenders of those who have given such conditions shall be treated as incomplete and will be summarily rejected.	Firm Delivery Schedule: All ordered goods shall be delivered within 40 days from the date of purchase order(s). Cross conditions such as "SUBJECT TO AVAILABILITY" "SUPPLIES WILL BE MADE AS AND WHEN SUPPLIES ARE RECEIVED" etc., will not be considered under any circumstances and the tenders of those who have given such conditions shall be treated as incomplete and will be summarily rejected.
	1.15.1	Delivery Period: The supply should be completed maximum within two weeks from the date of issue of order for Diphu and Dibrugarh and for rest of the location(s), supply is required within 6 weeks from the date of placement of the purchase order unless otherwise specified in the purchase order.	Delivery Period: The supply should be completed maximum within 40 days from the date of issue of purchase order unless otherwise specified in the purchase order.
	Annexure-VII and XII-A & B	List of item quoted and quantities in price bid	Quantity of each item shall be as mentioned above.
	Technical Specification	Technical Specifications of various equipment as given in the tender document.	Certain amendments are made in the technical specifications of various equipment. The same is attached as Annexure-A. Amended specs, if any, shall prevail.
	Time Schedule	<ul style="list-style-type: none"> <li>- Last date and time of bid submission (Online): 1<sup>st</sup> April'20 till 5:30 PM</li> <li>- Last date and time of submission of Bid (Technical) in hardcopy: 1<sup>st</sup> April 2020 till 05:30 PM</li> <li>- Bid Opening Date (online &amp; offline opening): 2<sup>nd</sup> April 2020 at 11:00 AM</li> </ul>	<ul style="list-style-type: none"> <li>- Last date and time of bid submission (Online): 23<sup>rd</sup> April till 5:30 PM</li> <li>- Last date and time of submission of Bid (Technical) in hardcopy: 23<sup>rd</sup> April 2020 till 05:30 PM</li> <li>- Bid Opening Date (online &amp; offline opening): 24<sup>th</sup> April 2020 at 11:00 AM</li> </ul>

Rest of tender terms and conditions remain same.

--sd--  
(Head Operations)

Amendment in Technical Specifications,				Annexure-A
Tender ref No: ACCF/MEDEQUIPMENT/19-20/01 for Medical Equipment				
Sl No.	Item No. and Name	Tender Specs/Clauses	Tender Technical Specs	Amendment made
1	Item no.1 Infusion Pump	Tech Specs	1. Should have the following flow rates. IV set ml/hr drops/min • 15 drops/ml 3-450 ml/hr 1-100 drops/min • 20 drops/ml 3-450 ml/hr 1-100 drops/min • 60 drops/ml 1-100 ml/hr 1-100 drops/min 2. Should have a flow rate accuracy of $\pm 10\%$ and drip rate accuracy of $\pm 2\%$ .	1. Should have the following flow rates. IV set ml/hr drops/min • 20 drops/ml 3-450 ml/hr 1-100 drops/min • 60 drops/ml 1-100 ml/hr 1-100 drops/min. 2. Should have a flow rate accuracy and drip rate accuracy of $\pm 5\%$ .

2		Tech Specs	<ul style="list-style-type: none"> <li>1. Should be operated on drip rate Peristaltic finger pump method</li> <li>2. Should have a flow rate accuracy of <math>\pm 10\%</math> and drip rate accuracy of <math>\pm 2\%</math></li> </ul>	<ul style="list-style-type: none"> <li>1. Should be operated on flow rate (ml/hr) mode</li> <li>2. Should have a flow rate accuracy of <math>\pm 5\%</math></li> </ul>
3	Item No. 2 Syringe Pump	Tech Specs	<ul style="list-style-type: none"> <li>1. Display of Drug Name with a provision of memorizing 10~15 names by the operator</li> <li>2. It should have Profile Function – Program infusion condition (flow rate, delivery volume, delivery time differently every hour, up to 24 hours</li> <li>3. It should having RS 232/HL 7 output for seamless ICU charting Integration</li> <li>4.3 Mounting device/ Docking Station for two or four pumps as per requirement so as to enable to power up to 2- 4 pumps with one power cord when mounted on IV pole. - 01( optional)</li> </ul>	<ul style="list-style-type: none"> <li>1. Display of Drug Library of more than 500 drugs.</li> <li>2. No Change</li> <li>3. No change</li> <li>4.3 Mounting device/ Docking Station for two or four pumps as per requirement and HL7 compliant so as to enable to power up to 2- 4 pumps with one power cord when mounted on IV pole. -01</li> </ul>

4		Tech Specs	Must Work on commonly available ISO/CE/FDA APPROVED/CERTIFIED 20, 50/60 ml Syringes with accuracy of minimum of +/-2% or better	Must Work on commonly available ISO/CE/FDA APPROVED/CERTIFIED 10ml,20ml, 50/60 ml Syringes with accuracy of minimum of +/-2% or better
5	Item No: 5. Defibrillator	Tech Specs	15. It should meet international safety standards, FDA and CE certification.	It should meet international safety standards, USFDA/European CE certification
6	Item No: 3. Patient Monitor	Tech Specs	1. The screen display size must be at least 10" with Colour TFT.2. It should have US-FDA Approval.	1. No Change.2. It should have European CE/ US-FDA Approval.
7	Item No: 7. Pulse oximeter	Tech Specs	1. Should have plethysmography wave form with numeric display for SPO2 and Heart rate on LCD/TFT display.	1. No change
8	Item No. 6: Portable ventilator	Tech Specs	1. Pressure Controlled mode 2. Shall have NIV in all modes 3. BIPAP/Bi-level/ASV/Equivalent 4. Tidal volume - (20– 1500)ml 5. Respiratory rate: 0-80 BPM 6. Oxygen Concentration - 21 – 100 %	1. No change 2. No change 3. No change 4. Tidal volume - (50– 1500) ml 5. No change 6. No change 7. No change.

			<p>7. Inspiratory Pressure - 4 – 50 cm H<sub>2</sub>O.</p> <p>8. Works with both high pressure and low pressure O<sub>2</sub>.</p> <p>9. Should have screen size 8 inch</p> <p>10. 220/240 V Ac 50 Hz supply. Internal battery (Li Ion) with (5-8) hours minimum operating time ( hot swappable allowed )</p> <p>11.Internal battery (maintenance free) with 5-8 hours minimum operating</p> <p>12. Works independent of gas cylinder pressure/compressor</p>	<p>8. No change</p> <p>9. No change</p> <p>10. No change</p> <p>11.No change</p> <p>12. No Change</p>
--	--	--	---	---