

Result of Technical Evaluation Report
for
Entering into Rate Contract for Supply of Medical Equipment at various locations of Assam

Date: 25.9.2020

1. Tender was floated for rate contracts of various medical equipment meant for Day Care Centers.
2. **Tender Reference No.:** ACCF/Medequip/19-20/01 dated 9.2.2020
3. **Mode of Bidding:** Two Cover Online System

The table below summarizes the list of bidders found qualified during evaluation of general and commercial requirements:

| Sr. No. | Brief Description of Items | Name of Bidder(s) | Qualification to General and commercial requirements | Reasons of disqualification |
|---------|----------------------------|----------------------------------|-----------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|
| 1 | Patient Monitor | Schiller India Pvt Ltd | Qualified | |
| | | Marak Supply & Agency | Qualified | |
| 2 | 12 channel ECG | Schiller India Pvt Ltd | Qualified for Model No. Cardio AT102 G2. Disqualified for Model No. Cardio AT2 Plus. | Cardio AT 2 plus is technically non-compliant as 12 leads ECG can not be viewed at a time (Point No. 10 of Technical Specs) |
| | | Marak Supply & Agency | Qualified | |
| | | Decent Medical Equipment, Indore | Disqualified | Manufacturer M/s Wuhan Zoncare China is not meeting desired average annual turnover criteria. |
| 3 | Defibrillator | Schiller India | Disqualified | Defiguard 400 is technically non-compliant on Tender Specs point 7 & 13 not being met. |
| 4 | Syringe Pump | Decent Medical Equipment, Indore | Qualified | |
| | | Akas Medical, Mumbai | Qualified | |

| | | | | |
|----|-------------------------------------|----------------------------------|--------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 5 | Infusion pump | Decent Medical Equipment, Indore | Qualified | |
| | | Akas Medical, Mumbai | Qualified | |
| 6 | Portable Ventilator | Decent Medical Supply | Disqualified | Technically Non-Compliant as ventilator doesn't meet many technical points as mentioned by the bidder itself in his Compliance sheet. Moreover, also doesn't meet FiO2 range requirements. |
| 7 | Pulse Oximeter | No quote | No quote | |
| 8 | Portable suction machine - high vac | No quote | No quote | |
| 9 | Vertical Autoclave | No quote | No quote | |
| 10 | Oxygen cylinders with flow meters | No quote | No quote | |
| 11 | Biosafety Cabinet | No quote | No quote | |

Financial bid of qualified bidders for Patient Monitors, ECG machine, Infusion Pump and Syringe Pumps shall be opened. Notification for the same shall be issued separately.

--sd--
(Finance Controller)
ACCF